

C,H, N elemental analysis

C, H and N were analyzed at Scientific and Technological Centers of the University of Barcelona using an elemental organic analyzer Thermo EA Flash 2000 (Thermo Scientific, Milan, Italy) working in standard conditions recommended by the supplier of the instrument (helium flow 140 ml/min, combustion furnace at 950 °C, chromatographic column oven at 65 °C).